

## PULSE TRANSIT TIME ESTIMATION FOR CONTINUOUS BLOOD PRESSURE MEASUREMENT: A COMPARATIVE STUDY

Diana Duarte<sup>I</sup>, N.V. Lopes<sup>I,II</sup>, Rui Fonseca-Pinto<sup>I,III</sup>

<sup>I</sup> School of Technology and Management. Polytechnic Institute of Leiria. Leiria, Portugal

<sup>II</sup> CITAB - Centre for the Research and Technology of Agro-Environmental and Biological Sciences. University of Trás-os-Montes and Alto Douro. Vila Real, Portugal

<sup>III</sup> Instituto de Telecomunicações. Leiria, Portugal.

**Introduction:** Continuous blood pressure (BP) monitoring provides important information about the cardiovascular system condition. Invasive methods are accurate but denote increased risk. Non-invasive methods are safe but less reliable and don't provide continuous information. An alternative approach for a continuous, non-invasive measurement of BP is based on changes in pulse transit time (PTT). PTT is defined as the time delay between the R-wave of the electrocardiogram (ECG) and the peak value of the photoplethysmogram (PPG) signal acquired in the patient finger on the same cardiac cycle.

**Objective:** The main goal of this work is to estimate the PTT using different methodologies found in the literature, such as derivative based approaches and Discrete Wavelet (DWT) and Hilbert-Huang (HHT) transforms, and compare the results with ground truth values obtained manually.

**Methods:** Several ECG and PPG signals were obtained from a public physiologic signals database and from real data using in vivo patients. A derivative based approach to find the R-waves and the PPG peak values was implemented to estimate the PTT. Then, a previous signal processing using DWT and HHT was applied. Results provided from each methodology were compared with ground truth values.

**Results and Conclusions:** Globally all methods had similar results when using the database signals. However, using real data, the derivative based approach performed poorly. Due to noise, non-stationary and non-linearity nature of physiologic signals, methodologies that incorporate DWT or HHT provide accurate results.

**Descriptors:** Blood Pressure, Pulse Transit Time, Wavelet Transform, Hilbert-Huang Transform.